

CERTIFICATE OF ANALYSIS

ERM[®]-BD475

Ochratoxin A (OTA) in ground roasted coffee		
Compound	Certified value ¹⁾	Uncertainty ²⁾
	Mass fraction in $\mu\text{g kg}^{-1}$	
Ochratoxin A	6.0	± 0.6

¹⁾ Unweighted mean of four independent results obtained by BAM using appropriate extraction, HPLC separation and MS/MS detection technique. The certified value is traceable to the SI.

²⁾ Estimated expanded uncertainty U with a coverage factor of $k = 2$, corresponding to a level of confidence of about 95 %, as defined in the Guide to the expression of uncertainty in measurement (GUM), ISO, 1995. Uncertainty contributions arising from the analytical method as well as from homogeneity and stability testing were taken into account.

The certified property value will be valid for 12 months beginning with the dispatch of the material from BAM; this validity may be extended as further evidence of stability becomes available.

The minimum sample intake is 6 g.

NOTE

European Reference Material ERM[®]-BD475 was produced and certified under the responsibility of BAM – Federal Institute for Materials Research and Testing according to the principles laid down in the technical guidelines of the European Reference Materials[®] co-operation agreement between BAM-LGC-IRMM. Information on these guidelines is available on the Internet (<http://www.erm-crm.org>).

Accepted as an ERM[®], Berlin, 05 October 2010

Date of dispatch:

Signed

BAM
Department I
Analytical Chemistry; Reference Materials
Berlin, Germany

BAM
Division I.2
Organic Chemical Analysis; Reference Materials
Berlin, Germany

Prof. Dr. U. Panne
(Head of Department)

Prof. Dr. I. Nehls
(Head of Division)

DESCRIPTION OF THE SAMPLE

The intended purpose of reference material ERM[®]-BD475 is i) validation of analytical procedures for the determination of ochratoxin A in food, and ii) quality assurance in the analytical laboratory.

The material ERM[®]-BD475 is a ground roasted coffee sample from commercial sources intended for human consumption and was spiked with OTA. After homogenisation, bottling was done using a version of the so-called “cross riffing” procedure. A total of 128 units were bottled in 250 mL amber glass bottles containing (64 ± 1) g sealed with screw caps containing PTFE-inlays, and numbered in the order of leaving the bottling process. The material is stored at BAM at -20 °C until dispatch.

The between-bottle homogeneity was evaluated by analysis of variance (ANOVA) on 8 out of 128 bottles (4 replicate analyses per bottle). Extensive stability tests provided sound evidence for a minimum validity of the certified value as indicated on page 1 of this certificate provided the material is stored according to the instructions given on page 3. Homogeneity and stability testing are described in detail in the certification report.

PARTICIPANTS

An interlaboratory comparison study (ILC) involving the following 16 laboratories was conducted in order to support the in-house certification study at BAM.

Laboratory	City / Country
Bayerisches Landesamt für Gesundheit und Lebensmittelsicherheit	Oberschleißheim, Germany
Bundesinstitut für Risikobewertung	Berlin, Germany
Chemisches und Veterinäruntersuchungsamt Sigmaringen	Sigmaringen, Germany
CR3-Kaffeeveredlung H. Hermsen GmbH	Bremen, Germany
General Chemical State Laboratory, Food Division & Division of Environment	Athens, Greece
Kantonales Labor Zürich	Zürich, Switzerland
Labor Eurofins Wiertz-Eggert-Jörissen	Hamburg, Germany
Landesuntersuchungsamt Rheinland-Pfalz	Trier, Germany
Landesuntersuchungsanstalt für das Gesundheits- und Veterinärwesen Sachsen	Dresden, Germany
Lebensmittelchemisches Institut Köln	Köln, Germany
Lebensmittelversuchsanstalt	Vienna, Austria
Max Rubner Institut	Detmold, Germany
National Food Administration Sweden	Uppsala, Sweden
Public Analyst's Laboratory Dublin	Dublin, Ireland
SGS Germany GmbH	Hamburg, Germany
Tchibo Manufacturing GmbH & Co KG	Hamburg, Germany

ANALYTICAL METHODS USED FOR CHARACTERISATION

For the in-house certification study at BAM an immunoaffinity clean-up step (IAC) after extraction was applied. High performance liquid chromatography (HPLC) technique was used for separation of the purified extract followed by MS/MS detection (MRM mode) with ¹³C₂₀ OTA as internal standard.

The ILC participant laboratories applied methods of their own choice. All laboratories used an IAC clean-up after extraction followed by HPLC analysis for separation of the purified extract. Fluorescence measurements were used for detection and quantification combined with an external calibration.

SAFETY INFORMATION

Ochratoxin A is a (suspected) carcinogen and should be handled with due caution. Although the OTA content in the samples is at trace levels, any use other than the intended of the content of the bottles is strictly prohibited. Personnel handling the material must adequately be trained and follow regular laboratory safety precautions.

INSTRUCTIONS FOR USE

Before withdrawing a sub-sample the bottle should be allowed to reach room temperature and mixed thoroughly. Thereafter, the bottle is to be closed tightly. To the best of our knowledge, the stability of the reference material is not affected by short periods of handling at ambient temperature during transport and use. However, BAM cannot be held responsible for any alterations of the material occurring during transportation to, and handling and storage at, the customer's premises, especially of opened samples.

STORAGE

The material has to be stored at a temperature equal to or lower than 20 °C in its original bottle.

LEGAL NOTICE

Neither BAM, its contractors nor any person acting on their behalf:

- (a) make any warranty or representation, express or implied, that the use of any information, material, apparatus, method or process disclosed in this document does not infringe any privately owned intellectual property rights; or
- (b) assume any liability with respect to, or for damages resulting from, the use of any information, material, apparatus, method or process disclosed in this document save for loss or damage arising solely and directly from the negligence of BAM.

TECHNICAL REPORT

A detailed technical report (pdf file or paper copy; in English) describing the production, general characterisation as well as the analytical procedures applied and the treatment of the analytical data during certification of ERM[®]-BD475 is available on request from BAM.

Supply of Reference Materials by BAM Federal Institute for Materials Research and Testing

Richard-Willstätter-Straße 11, D-12489 Berlin, Germany

Phone: +49 30 8104 1129

e-mail: sales.crm@bam.de

Fax: +49 30 8104 1127

internet: www.bam.de